



ASEAN Reference Substance (ARS)

ASEAN TECHNICAL COOPERATION IN PHARMACEUTICALS
supported by Japan Pharmaceutical Manufacturers Association



Coordinator of the Project : Bureau of Drug and Narcotic, Department of Medical Sciences, Ministry of Public Health
Nonthaburi 11000, Thailand TEL.(662)9510000 ext. 99102, 99103 FAX.(662)5805733

Coordinating Country : Centre for Quality Control, National Pharmaceutical Control Bureau, Ministry of Health,
Jalan University P.O.Box 319, 46730 Petaling Jaya, Malaysia
TEL.(603) 78835400, 78835410 FAX (603) 79562924

Certificate of Analysis

CIMETIDINE

Control No. M 213066

Description : A white, crystalline powder

Infrared absorption : Concordant with the reference spectrum of Cimetidine USPRS

Light absorption : A solution (1 in 80,000) in 0.1 N sulfuric acid exhibits maxima and minima at the same wavelengths as that of a similar solution of USPRS

Melting range : 141 °C

Chromatographic purity : (HPLC)

Any single impurity : Not more than 0.2%

Total impurities : Not more than 1.0%

Loss on drying : 0.1%

Assay : 100.02% of $C_{10}H_{16}N_6S$, calculated on the dried basis, determined by HPLC method, compared with USPRS

Intended use : For HPLC, chemical assay and identification

Direction for use : Dry at 110°C for 2 hours before use

Storage : Keep container tightly closed and protected from light, preferably at the temperature 2 - 8 °C

Date of Adoption : 25 February 2013

Retested Date : 21 March 2016, 1 October 2019

Next Retest Date : 1 October 2022